

Declaration of Conformity

Manufacturers Name: HTZ Limited
Manufacturers Address: Vulcan Way
New Addington
Croydon
CR0 9UG
United Kingdom
Tel: +44 (0) 1689 843345
Fax: +44 (0) 1689 841792

Type of Equipment: Agglutination Reader
Model Number: Duet

This is to certify that the above instrument conforms with the essential Requirements of the following Council Directives:-

- Electromagnetic Compatability Directive 89/336/EEC,
- Low Voltage Directive 73/23/EEC
- In Vitro Device Directive 98/79/EEC (Annex III sections 1 - 5 & its amending directives)


The following EMC standards have been applied during testing:

- | | |
|--|---------------|
| • Harmonic | EN 61000-3-2 |
| • Flicker | EN 61000-3-3 |
| • Immunity Test - ESD | EN 61000-4-2 |
| • Radiated Immunity test | EN 61000-4-3 |
| • Electrical fast transient/Burst immunity test | EN 61000-4-4 |
| • Immunity Test - Conducted RF | EN 61000-4-6 |
| • Immunity Test - Voltage Surges (AC Power Port) | EN 61000-4-5 |
| • Immunity Test - Voltage Dips & Interruptions | EN 61000-4-11 |
| • Conducted & Radiated Emissions | EN 55022 |

The following product safety standards have been applied during testing:

- | | |
|-------------------------------|---------------------|
| • General Safety requirements | EN 61010-1:2001 |
| • IVD Safety requirements | EN 61010-2-101:2002 |

Signed:



Full Name: Neil Woodmansey

Position: Company Director

Date: May 2007